

University of Connecticut
Daily Temperature Log for Warmers

Month / Year _____

Clinical Unit _____

			Notification & Problem Resolution Comments	
	*TEMP	INITIALS		
1			___ Fac. Mgmt.	Resolution:
2			___ Fac. Mgmt.	Resolution:
3			___ Fac. Mgmt.	Resolution:
4			___ Fac. Mgmt.	Resolution:
5			___ Fac. Mgmt.	Resolution:
6			___ Fac. Mgmt.	Resolution:
7			___ Fac. Mgmt.	Resolution:
8			___ Fac. Mgmt.	Resolution:
9			___ Fac. Mgmt.	Resolution:
10			___ Fac. Mgmt.	Resolution:
11			___ Fac. Mgmt.	Resolution:
12			___ Fac. Mgmt.	Resolution:
13			___ Fac. Mgmt.	Resolution:
14			___ Fac. Mgmt.	Resolution:
15			___ Fac. Mgmt.	Resolution:
16			___ Fac. Mgmt.	Resolution:
17			___ Fac. Mgmt.	Resolution:
18			___ Fac. Mgmt.	Resolution:
19			___ Fac. Mgmt.	Resolution:
20			___ Fac. Mgmt.	Resolution:
21			___ Fac. Mgmt.	Resolution:
22			___ Fac. Mgmt.	Resolution:
23			___ Fac. Mgmt.	Resolution:
24			___ Fac. Mgmt.	Resolution:
25			___ Fac. Mgmt.	Resolution:
26			___ Fac. Mgmt.	Resolution:
27			___ Fac. Mgmt.	Resolution:
28			___ Fac. Mgmt.	Resolution:
29			___ Fac. Mgmt.	Resolution:
30			___ Fac. Mgmt.	Resolution:
31			___ Fac. Mgmt.	Resolution:

Per hospital policy, warmers will not be set at a temperature higher than 130°F for blankets and 110°F for fluids. If the warmer exceeds these temperatures it must be readjusted downward or a call must be placed to Facilities at x2125 for repair, as appropriate, and document your actions in the follow-up column.

This record must be kept for one year and then destroyed per State requirement.

*** If the warmer temperature falls out of range please document your actions in the follow-up column.**